

DBH
TIN

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 1/15/2015**

Instrument Serial Number: 010557

As of 15-Jan-15

Start Date	End Date	Agency Name
13-Jan-15		DFS Central Lab
05-Aug-14	13-Jan-15	Amelia County SO
24-Jun-14	05-Aug-14	DFS Central Lab
07-Jan-09	24-Jun-14	Richmond County Sheriff's Office
15-Aug-08	07-Jan-09	DFS Central Lab

DBH
TLW**INTOX EC/IR II**
Quality Assurance Worksheet

Instrument Serial Number 010557 Worksheet Start Date 1/13/2015

Location Amelia CO SO

Address 16441 Court St Amelia VA 23002

DFS Technician Donald Hall License No. 21250

☐ Laboratory☒ On-SiteSite Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 760 Reference Barometer (mm HG) 763

Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.301
0.300	0.300	0.291	0.309	Sample 2	0.301
Precision		sample min	sample max	Sample 3	0.301
0		0.301	0.301		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		

Dry gas standard Lot No. (with tank no.)

AG332301-06

☐ Replaced dry gas standard (+O-ring)☐ Installed at Location☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010557

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

1/15/15

Issuing Analyst

[Signature]

Date

*1/21/15**JPD*

DBH
TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 01/13/2015
TEST EQUIPMENT NUMBER 010557		

RESULTS: TIME SAMPLE TAKEN 14:12 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 13 Jan 2015 at 14:37

DBH
TLW

Test Results

Instrument Serial Number 010557

Test # 000357 Subject Test

Test Location 1 Department of
Test Date 13 Jan 2015

Test Location 2 Forensic Science
Test Time 14:06
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 13 Jan 2015 End Time 14:13

Result Time 14:12

Result Date 13 Jan 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:06

Data Type BLK

Sample Value 0.000

Sample Time 14:07

Data Type CHK

Sample Value 0.100

Sample Time 14:07

Data Type BLK

Sample Value 0.000

Sample Time 14:09

Data Type SUBJ

Sample Value 0.000

Sample Time 14:09

Data Type BLK

Sample Value 0.000

Sample Time 14:10

Data Type SUBJ

Sample Value 0.000

Sample Time 14:12

Data Type BLK

Sample Value 0.000

Sample Time 14:13

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG332301-06

Standard Expiration Date 11/19/2015

Tank Pressure 1051

Barometric Pressure 760 mmHg

Blow Sample Number 1 Blow Duration 3.34 sec

Blow Volume 1748 cc End-of-Blow Time 14:09

Blow Sample Number 2 Blow Duration 3.85 sec

Blow Volume 2056 cc End-of-Blow Time 14:12

Tamper Evident Stamp a2cfe9d8

Test Status Code 0

Test Status Success

Intox EC/1R-II: Accuracy Check

DBH
TLN

Department of Forensic Science

Serial Number: 010557 Test Number: 358

Test Date: 01/13/2015 Test Time: 14:15 EST

Dry Gas Target: 0.100

Lot Number: AG332301-06 Exp Date: 11/19/2015

Tank Pressure: 1046 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:16
CHK	0.100	14:17
BLK	0.000	14:18
CHK	0.100	14:19
BLK	0.000	14:20
CHK	0.100	14:21

Test Status: *Success*

Calibration CRC: 3E7EEA90

Intox EC/IR-II: Accuracy Check

DBH
TLW

Department of Forensic Science

Serial Number: 010557 Test Number: 359

Test Date: 01/13/2015 Test Time: 14:24 EST

Dry Gas Target: 0.300

Lot Number: AG428003-01 Exp Date: 10/07/2016

Tank Pressure: 237 psi Barometric Pressure: 760 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	14:24
CHK	0.301	14:25
BLK	0.000	14:27
CHK	0.301	14:28
BLK	0.000	14:29
CHK	0.301	14:30

Test Status: Success

Calibration CRC: 3E7EEA90